



Product Service

Attestation of Conformity


No. N8 13 03 84161 002

Holder of Certificate: UREATac Co., Ltd.974-7, Yangdeok-dong, Masan Hoewon-gu,
Gyeongsangnam-do,
Changwon-si, 630-490
REPUBLIC OF KOREA**Product:** Ventilation Equipment
(Individual Ventilation Cage system)**Model(s):** ACU-M-V01, ACU-R-V01**Parameters:**

Rated input voltage :	100-240 Va.c., 50/60 Hz
Rated input current :	4.2 A
Protection class :	I
Degree of protection against ingress of liquids :	IPX0

Tested according to: EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: MAEB0105693
(James Jeon)**Date,** 2013-03-27

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1